

Intox EC/IR-II: Subject Test

Syracuse Police Dept. 310 N Huntington St. Syracuse, IN 46567

Serial Number: 011037 Test Number: 569
Test Date: 08/03/2024 Test Time: 14:07 EDT

Operator Name: Kenworthy, Corey R
Operator Certification Number: [REDACTED]
Agency Name: Syracuse Police
Observation Began: 08/03/2024 at 13:54
Observer Name: Kenworthy, Corey, R
Driver License Number: [REDACTED]
Subject Name: MAGLEY, ROBERT E
Subject D.O.B.: [REDACTED]

Dry Gas Target: 0.078
Lot Number: AG318012-3WEN6 Tank Number: 5 Exp Date: 06/29/2025

System Check: Passed

Test	g/210L	Time
BLK	0.000	14:09
CHK	0.077	14:10
BLK	0.000	14:11
SUBJ	0.156	14:13
BLK	0.000	14:15
SUBJ	0.149	14:17
BLK	0.000	14:19
CHK	0.077	14:20
BLK	0.000	14:21

Test Status Sample Complete

RESULT: 0.149 g/210L
14:17 EDT,
08/03/2024

ALCOHOL READINGS ARE EXPRESSED AS
GRAMS OF ALCOHOL PER 210 LITERS OF BREATH

[REDACTED]

Operator Signature